



Republic of the Philippines  
Department of Agriculture

**SUGAR REGULATORY ADMINISTRATION**


Sugar Center Bldg., North Avenue, Diliman, Quezon City, Philippines 1101

TIN 000-784-336

Website: <http://www.sra.gov.ph>

Email Address: [srahead@sra.gov.ph](mailto:srahead@sra.gov.ph)

Tel No. (632) 8929-3633, (632) 3455-2135, (632) 3455-3376



**SUPPLEMENTAL/BID BULLETIN NO. 1**  
**Supply and Delivery of Two (2) Units Electric Autoclave**  
**Bid Reference No. 2023-SIDA-12**

With reference to the Bidding Documents for the **Supply and Delivery of Two (2) Units Electric Autoclave**, posted at the PhilGEPS and SRA websites, this Supplemental/Bid Bulletin is being issued to clarify, amend, modify some items in the provisions of the Bidding Documents in response to the recommendations raised by prospective bidders during the pre-bid conference and in writing, as follows:

Section VII. Technical Specifications		
From	To	
(none stated)	A.24	Capacity: 50L minimum

Response to Prospective Bidder's Queries

SECTION VI. SCHEDULE OF REQUIREMENTS				
BIDDERS	ITEM NO.	SRA SPECIFICATION	RECOMMENDATIONS	REPLY/COMMENT
Prince Valiant Internal Corporation		Delivery Schedule: 60 calendar days upon receipt of the Notice to Proceed	Request for delivery extension up to <b>90 calendar days</b>	Denied.
MMJ Biosystems, Inc.		Delivery Schedule 60 calendar days upon receipt of the Notice to Proceed	Request for delivery extension up to <b>90 calendar days</b>	Denied.
SECTION VII. TECHNICAL SPECIFICATIONS				
BIDDERS	ITEM NO.	SRA SPECIFICATION	RECOMMENDATIONS	REPLY/COMMENT
Prince Valiant Internal Corporation		Capacity: <i>None</i>	Capacity: 50 - 60 L	50L minimum
	A.2	Operating Range: within 105°C to 135°C	Operating Range: within 105°C to 140°C	Denied. Retain original specifications
	A.23	Consumption: 2000 watts (20000 VA)	Consumption: 2000 - 3000 watts	Denied. Retain original specifications
MMJ Biosystems, Inc.		Capacity: <i>None</i>	Inquiry as to Autoclave capacity	50L minimum
			Inquiry as to whether the bidder can make a counter-offer of better specifications or within the range of the parameters of autoclave	"Better specifications" should be stated.



Management System  
ISO 9001:2015  
[www.tuv.com](http://www.tuv.com)  
ID 910508567



SECTION VII. TECHNICAL SPECIFICATIONS				
BIDDERS	ITEM NO.	SRA SPECIFICATION	RECOMMENDATIONS	REPLY/COMMENT
Levins International	A.4	Warming Range: within 45°C to 60°C	Warming Range: within 45°C to 95°C programmable	Denied. Retain original specifications
	A.9	Maximum Allowable: 38 psi (0.26MPa)	Maximum Allowable: 0.263 Mpa	This is also equivalent to 38 psi
	A.12	1 to 250 minutes	1-99 hours (selectable)	Denied. Retain original specifications
		Volume: <i>None</i>	Internal volume: 58 L	50L minimum
			Floor standing top open lid, foot pedal to open lid	
	A.13	Delayed Auto-start	Operation start timer	Denied. Retain original specifications
	A.15	<b>Dual-sensing lid interlock</b> , over-pressure power cutoff, over temperature power cutoff, low-water power cutoff, mechanical pressure relief valve, microprocessor-controlled pressure relief valve, lid closure detector, over-current detector, and circuit breaker	Safety device lid interlock mechanism	Denied. Retain original specifications
RainPhil, Inc.		Volume: <i>None</i>	Inquiry as to volume or capacity of Autoclave	50L minimum

The Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified and superseded accordingly.

For the guidance and information of all concerned.

Issued this 25<sup>th</sup> day of August 2023.

  
**ATTY. IGNACIO S. SANTILLANA**  
Chairperson, Regional Bids and Awards Committee