



Republic of the Philippines
 Department of Agriculture
SUGAR REGULATORY ADMINISTRATION
 Sugar Center Bldg., North Avenue, Diliman, Quezon City, Philippines 1101
 TIN 000-784-336
 Website: <http://www.sra.gov.ph>
 Email Address: srahead@sra.gov.ph
 Tel. No.: (632)8929-3633, (632)3455-2135, (632)3455-3376

SUPPLEMENTAL/BID BULLETIN NO. 1
Supply and Delivery of Laboratory Equipment
ITB No. 2021-SIDA-02

With reference to the Bidding Documents for the supply and delivery of laboratory equipment, posted at the PhilGEPS and SRA websites, this Supplemental/Bid Bulletin is being issued to clarify, amend, modify the provisions in the Bidding Documents in response to the queries and recommendations raised by prospective bidders, as follows:

Section I. Invitation to Bid Par. 5	
From	To
<p>2. The Sugar Regulatory Administration now invites bids for Supply and Delivery of Laboratory Equipment. Delivery of the Goods is required by November 9, 2021. Bidders should have completed, within five (5) years from the date of submission and receipt of bids, a contract similar to the Project. The description of an eligible bidder is contained in the Bidding Documents, particularly, in Section II (Instructions to Bidders).</p>	<p>2. The Sugar Regulatory Administration now invites bids for Supply and Delivery of Laboratory Equipment. Delivery of the Goods is required by December 15, 2021. Requests for delivery extensions will not be accepted. Bidders should have completed, within five (5) years from the date of submission and receipt of bids, a contract similar to the Project. The description of an eligible bidder is contained in the Bidding Documents, particularly, in Section II (Instructions to Bidders).</p>
<p>5. A complete set of Bidding Documents may be acquired by interested Bidders on August 17, 2021 from the given address and website below and upon payment of the applicable fee for the Bidding Documents, pursuant to the latest Guidelines issued by the GPPB, in the amount of PhP 6,500. The Procuring Entity shall allow the bidder to present its proof of payment for the fees in person.</p>	<p>5. A complete set of Bidding Documents may be acquired by interested Bidders on August 17, 2021 from the given address and website below and upon payment of the applicable fees for the Bidding Documents, pursuant to the latest Guidelines issued by the GPPB, in the amounts stated hereunder. The Procuring Entity shall allow the bidder to present its proof of payment for the fees in person. The Procuring Entity shall also allow the bidder/s to deposit payment for the bid document fees on SRA's Landbank Current Account: (Account Name: SUGAR REGULATORY ADMINISTRATION Account Number: 0422-1062-51 LBP Branch: Gatuslao Branch)</p>



Management System
 ISO 9001:2015
 www.tuv.com
 ID 9105085567



A food-secure and resilient Philippines
 with empowered and prosperous farmers and fisherfolk



Lot	Description	Bid Documents Fee (₱)
1	Chemical Cabinets, 1 set	500.00
2	Leaf Grinder, 1 unit	500.00
3	Microbalance, 1 unit	1,000.00
4	Microwave Digestor, 1 unit	3,000.00
5	Water purification system (whole assembly), 1 unit	1,500.00
	TOTAL	6,500.00

Section III. Bid Data Sheet Clause 14.1																																												
From	To																																											
<p>The bid security shall be in the form of a Bid Securing Declaration, or any of the following forms and amounts:</p> <p>a. The amount of not less than PhP 130,000.00, if bid security is in cash, cashier's/manager's check, bank draft/guarantee or irrevocable letter of credit; or</p> <p>b. The amount of not less than PhP 325,000.00, if bid security is in Surety Bond.</p>	<p>The bid security shall be in the form of a Bid Securing Declaration, or any of the following forms and amounts:</p> <p>a. The amount of not less than the indicated amounts below, if bid security is in cash, cashier's/manager's check, bank draft/guarantee or irrevocable letter of credit; or</p> <table border="1"> <thead> <tr> <th>Lot</th> <th>Description</th> <th>ABC (₱)</th> </tr> </thead> <tbody> <tr> <td>1</td> <td>Chemical Cabinets, 1 set</td> <td>10,000.00</td> </tr> <tr> <td>2</td> <td>Leaf Grinder, 1 unit</td> <td>10,000.00</td> </tr> <tr> <td>3</td> <td>Microbalance, 1 unit</td> <td>20,000.00</td> </tr> <tr> <td>4</td> <td>Microwave Digestor, 1 unit</td> <td>60,000.00</td> </tr> <tr> <td>5</td> <td>Water purification system (whole assembly), 1 unit</td> <td>30,000.00</td> </tr> <tr> <td></td> <td>TOTAL</td> <td>130,000.00</td> </tr> </tbody> </table> <p>b. The amount of not less than the indicated amounts below if bid security is in Surety Bond.</p> <table border="1"> <thead> <tr> <th>Lot</th> <th>Description</th> <th>ABC (₱)</th> </tr> </thead> <tbody> <tr> <td>1</td> <td>Chemical Cabinets, 1 set</td> <td>25,000.00</td> </tr> <tr> <td>2</td> <td>Leaf Grinder, 1 unit</td> <td>25,000.00</td> </tr> <tr> <td>3</td> <td>Microbalance, 1 unit</td> <td>50,000.00</td> </tr> <tr> <td>4</td> <td>Microwave Digestor, 1 unit</td> <td>150,000.00</td> </tr> <tr> <td>5</td> <td>Water purification system (whole assembly), 1 unit</td> <td>75,000.00</td> </tr> <tr> <td></td> <td>TOTAL</td> <td>325,000.00</td> </tr> </tbody> </table>		Lot	Description	ABC (₱)	1	Chemical Cabinets, 1 set	10,000.00	2	Leaf Grinder, 1 unit	10,000.00	3	Microbalance, 1 unit	20,000.00	4	Microwave Digestor, 1 unit	60,000.00	5	Water purification system (whole assembly), 1 unit	30,000.00		TOTAL	130,000.00	Lot	Description	ABC (₱)	1	Chemical Cabinets, 1 set	25,000.00	2	Leaf Grinder, 1 unit	25,000.00	3	Microbalance, 1 unit	50,000.00	4	Microwave Digestor, 1 unit	150,000.00	5	Water purification system (whole assembly), 1 unit	75,000.00		TOTAL	325,000.00
	Lot	Description	ABC (₱)																																									
1	Chemical Cabinets, 1 set	10,000.00																																										
2	Leaf Grinder, 1 unit	10,000.00																																										
3	Microbalance, 1 unit	20,000.00																																										
4	Microwave Digestor, 1 unit	60,000.00																																										
5	Water purification system (whole assembly), 1 unit	30,000.00																																										
	TOTAL	130,000.00																																										
Lot	Description	ABC (₱)																																										
1	Chemical Cabinets, 1 set	25,000.00																																										
2	Leaf Grinder, 1 unit	25,000.00																																										
3	Microbalance, 1 unit	50,000.00																																										
4	Microwave Digestor, 1 unit	150,000.00																																										
5	Water purification system (whole assembly), 1 unit	75,000.00																																										
	TOTAL	325,000.00																																										

Section VII. Technical Specifications			
From		To	
C	Microbalance, 1 unit	C	Microbalance, 1 unit
C.14	Professional installation	C.14	Capacity: 5 g
C.15	Operations training and trouble shooting	C.15	Resolution: 0.000001 g
C.16	With compatible AVR for the unit	C.16	Professional installation
C.17	Warranty	C.17	Operations training and trouble shooting
		C.18	With compatible AVR for the unit
		C.19	Warranty
D	Microwave Digestor, 1 unit	D	Microwave Digestor, 1 unit
D.1	Stainless steel housing with multilayer PTFE coating	D.1	Stainless steel housing with multilayer PTFE or Teflon or PFA coating
D.5	Top loading or front loading, with vent system	D.5	Top loading or front loading, with vent system (removed)
D.10	Can accommodate: 8-12 vessels, and digest sugarcane leaf samples	D.10	Can accommodate: 8-20 vessels, and digest sugarcane leaf samples
D.12	With AVR and power supply compatible with the equipment	D.12	With AVR and UPS compatible with the equipment
D.14	Training and trouble shooting	D.14	Comprehensive and intensive training, troubleshooting, and preventive maintenance (PM) of the whole equipment assembly
D.19	No venting into the oven chamber and no pressure releasing door	D.19	No venting into the oven chamber and no pressure releasing door-(removed)
E	Water purification system (whole assembly)	E	Water purification system (whole assembly)
E.1	E.1 AC/DC input: 100V/240 V AC, 50-60 H	E.1	E.1 AC/DC input: 90V -240V AC, 50-60H
E.3	12-15L/hr flowrate	E.3	10 -15L/hr flowrate; 30 L/hr flowrate for type 2
E.4	25-40 kg operation weight	E.4	25- 150 kg operation weight
E.5	290-300 L/day max. usage	E.5	200 -300 L/day max. usage
E.6	Up to 2L/min distribution flow rate (from POD)	E.6	1 to 2L/min distribution flow rate (from POD)
E.7	Input: potable tap water	E.7	Input: potable tap water or softened and hardness stabilized
E.11	With storage tank and remote dispenser	E.11	Capable in producing both type 1 and 2 water
E.12	Professional installation	E.12	With a pre-treatment system (series of highly filtered series):
E.13	Operations training and trouble shooting	E.13	Equipped with Ultra-filtration (UF) and ultraviolet (UV):
E.14	With compatible AVR for the unit	E.14	Consumables included especially the filters:
E.15	Warranty	E.15	1-2 gallons per hour production rate capacity.
		E.16	With 60-L storage tank and remote dispenser

		E.17	Professional installation
		E.18	Operations training and trouble shooting
		E.19	With compatible AVR for the unit
		E.20	Warranty

CLARIFICATIONS ON PROSPECTIVE BIDDER'S WRITTEN QUERIES

BIDDER	QUERY	REPLY
RAINPHIL, INC.	1. Item B (Leaf Grinder): Is counteroffer allowed? (Provided that the item that is going to be offered is a higher profile than what is posted in Philgeps)	“The specifications and other terms in the bidding documents reflect minimum requirements. A bidder may be allowed to submit a superior offer, however , in the evaluation of the technical proposals – a “pass/fail” method - such that the presence or absence of the technical requirements is the sole basis for determining technical compliance. After having established compliance with the technical specifications, the next factor to consider would then be the price or financial bid.” (GPPB GPM Vol. 2)
	2. Is it Original, Copy 1, 2 and 3 submissions?	Yes. One (1) Original Copy and three (3) additional copies for the first and second component of the bids must be submitted.
	3. Do you accept a letter of delivery extension as long as it is not later than December 31, 2021?	Please refer to highlighted amendment under Section I. Invitation to Bid Par. 2 of this Supplemental/Bid Bulletin.
	4. E.4: What is the significance of this weight? Is this the weight of the reservoir? (Reservoirs of different capacities have different weights: with water or no water) Can it possibly be deleted?	DENIED. Specification remains the same.
	5. E.5: These values are mainly dependent from your water supply and that is the supply from the faucet. You may or you may not able to achieve your desired flow rate if the pressure of the feed water supply is low. Water Purification System always has feedwater requirements. Can it possibly be deleted?	DENIED. Specification remains the same.

	6. E.10: This is Type 1 Conductivity....the lesser the value, the better.	DENIED. Specification remains the same.
GUILBERN CORPORATION	1. Just want to clarify regarding for the payment of Bid Documents, we will pay for the per lot items or we will pay the amount of 6,500.00?	Please refer to the amendment under Section I. Invitation to Bid Par. 5 of this Supplemental/Bid Bulletin.
MOLAVE CORPORATION	1. E.1 AC/DC input: 90 to 240V 50/60Hz	GRANTED. Please refer to the highlighted amendment under Section VII. Technical Specifications E.1 of this Supplemental/Bid Bulletin.
	2. E.3. 30L/hr flow rate (Type 2)	GRANTED. Acceptable Please refer to the highlighted amendment under Section VII. Technical Specifications E.3 of this Supplemental/Bid Bulletin.
	3. E.4. 66 kg/ 146 lbs product wt.	GRANTED. Within range. Please refer to the highlighted amendment under Section VII. Technical Specifications E.4 of this Supplemental/Bid Bulletin.
	4. E.5. 300L/ day usage	Within the specified range.
	5. E.6. 1.5 L/min flowrate (Type 1) from POD	Within the specified range.
	6. E.7.Potable water or softened and hardness stabilized	GRANTED. Please refer to the highlighted amendment under Section VII. Technical Specifications E.7 of this Supplemental/Bid Bulletin.
	7. E.9.Type 1 water for: Cell and tissue culture PCR, DNA sequencing , Electrophoresis, TOC Measurements, IC , HPLC, GC-MS, ICP-MS, AA Type 2 water for: Rinsing lab glassware Supplying general lab equipment Preparing and diluting buffers, reagents, and media	Specifications remain the same. Bidder's offer will be subject to post-qualification.
	8. E.10. Type 1: 0.055µS/ Type 2: 0.1 to 1.0µS/cm	Specifications remain the same. Bidder's offer will be subject to post-qualification.
	9. E.11.With 100L integrated reservoir and remote dispenser	Specifications remain the same. Bidder's offer will be subject to post-qualification.

The Supplemental/Bid Bulletin No. 1 shall form part of the Bidding Documents. Any provisions in the Bidding Documents inconsistent herewith is hereby amended, modified, and superseded accordingly.

For the guidance and information of all concerned.

Issued this 28th day of August 2021.



ATTY. IGNACIO S. SANTILLANA
Chairperson
Regional Bids and Awards Committee